



900 42nd Street South
PO Box 6757
Fargo, ND 58108-6757

September 29, 2008

GF Health Products
Attention: Beatrice Sherer
2935 Northeast Parkway
Atlanta Georgia 3360

PDAC
Medicare Pricing, Data Analysis
and Coding

Re: John Bunn O2 Concentrator Model Number JB0180, JB0180S
Xref #: 6445421

Dear Ms. Beatrice Sherer:

This letter is in response to your recent inquiry for coding verification of the above listed product(s) manufactured and/or distributed by your company.

It is our determination that the Medicare HCPCS code(s) to bill the four DME Medicare Administrative Contractors (DME MACs) is/are:

E1390 OXYGEN CONCENTRATOR, SINGLE DELIVERY PORT, CAPABLE OF DELIVERING 85 PERCENT OR GREATER OXYGEN CONCENTRATION AT THE PRESCRIBED FLOW RATE

This HCPCS coding decision applies to the submitted product(s) as presented to and reviewed by the Pricing, Data Analysis and Coding (PDAC). Any modifications to the product(s) could change the HCPCS code and would need to be reviewed for coding verification. The assignment of a HCPCS code to the product(s) should in no way be construed as an approval or endorsement of the product(s) by PDAC or Medicare, nor does it imply or guarantee claim reimbursement or coverage. For questions regarding claim coverage or reimbursement, please contact your DME MAC jurisdiction.

Should you disagree with this coding decision, a reconsideration can be requested. The reconsideration form is located on the PDAC website (www.dmepdac.com) under the HCPCS Review tab. The PDAC will reconsider the request if made within 45 days of the date of this letter and additional supporting documentation is provided. If a request for





PDAC
Medicare Pricing, Data Analysis
and Coding

900 42nd Street South
PO Box 6757
Fargo, ND 58108-6757

reconsideration is made after 45 days, the request is treated as a new Coding Verification Review and a complete application must be submitted, along with the additional documentation supporting the request.

Should you have questions, regarding this decision, you may contact the PDAC Contact Center at the address listed above or by telephone at (877) 735-1326.

Sincerely,

Deb Kerlin, RN,BSN,CPC

Deb Kerlin, RN,BSN,CPC
PDAC Coding Analyst